

**Statistical Analysis of Organic Carbon (OC), Elemental Carbon (EC)  
& Total Carbon (TC) Concentrations Collected on Filters from the  
CEN TC 264 WG 35 Field Validation Exercise**

**D Butterfield, S Beccaceci & P Harris**

**MAY 2015**



Statistical Analysis of Organic Carbon (OC), Elemental Carbon (EC) &  
Total Carbon (TC) Concentrations Collected on Filters from the CEN TC  
264 WG 35 Field Validation Exercise

D Butterfield, S Beccaceci & P Harris

Analytical Science Division

© NPL Management Limited, 2015

ISSN 1754-2928

National Physical Laboratory  
Hampton Road, Teddington, Middlesex, TW11 0LW

Extracts from this report may be reproduced provided the source is acknowledged  
and the extract is not taken out of context.

Approved on behalf of NPLML by Dr Martyn Sene, Operation Director

## Executive Summary

Daily PM<sub>2.5</sub> samples were collected from 6 monitoring stations over 2013 & 2014 and analysed by 4 different laboratories. Sub-samples of each daily filter were analysed twice using 6 thermal-optical protocols (EUSAAR2, NIOSH & IMPROVE) with optical corrections for OCEC split point performed by both reflectance and transmittance methods. This resulted in 12 results per laboratory, per sampled filter.

NPL performed a statistical analysis to determine the uncertainty of a single measured value based on between-laboratory variability, internal-laboratory variability and between sampler variability. NPL expanded this analysis to rank the different protocol / optical correction methodologies on a site specific and complete dataset basis.

The analysis shows that the reflectance protocols have problems detecting very low levels of elemental carbon and are unable to deliver the detection limits required.

The results from the transmittance protocols show no protocol clearly performing better than the others across all of the site types and concentrations encountered during the field validation tests. EUSAAR2 performs best for EC at urban background and roadside locations and second best at rural locations.

There was no significant between sampler variability detected for most components and most protocols. Where it was detected it was insignificant compared to the internal repeatability.

The uncertainty of a single measurement result (95%, 2 $\sigma$ ) for the transmittance protocols is less than 25% above concentrations of 2  $\mu\text{g.cm}^{-2}$ . For concentrations below this the relative uncertainty is greater than 25%, but probably limited by the detection limit. When performing repeat measurements of a single sampled filter the repeatability was similar to the calculated uncertainty of a single measurement result.

The OC detection limit measured from laboratory blank filters was close to or exceeded the requirements of the draft standard, with the NIOSH protocols over reporting OC levels compared to the other protocols. The mean EC detection limit for reflectance protocols is  $<0.1 \mu\text{g.cm}^{-2}$ , while the mean detection limit for transmittance protocols was  $\sim 0.4 \mu\text{g.cm}^{-2}$ . This was above the proposals for the draft standard and is higher than that seen in other measurement studies.

Field blank OC and EC content was at or below the detection limit results calculated from the laboratory blanks.

At all the field validation sites two different types of automatic optical black carbon analysers (Multi Angle Absorption Photometer & Aethalometer) were co-located with the filter samplers. Results from these automatic analysers were compared against the filter measurements using the EU Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods and the EU Excel spreadsheet.

## CONTENTS

|                                                                                               |           |
|-----------------------------------------------------------------------------------------------|-----------|
| <b>1. INTRODUCTION .....</b>                                                                  | <b>1</b>  |
| <b>2. ANALYSIS METHODOLOGY .....</b>                                                          | <b>1</b>  |
| 2.1 CALCULATING BETWEEN- AND WITHIN-LABORATORY VARIABILITY .....                              | 1         |
| 2.1.1 Data processing.....                                                                    | 2         |
| 2.1.2 Outlier rejection.....                                                                  | 2         |
| 2.1.3 Data normalisation.....                                                                 | 3         |
| 2.1.4 Analysis of variance .....                                                              | 3         |
| 2.1.5 Calculation of standard deviations.....                                                 | 4         |
| 2.2 CALCULATING BETWEEN-SAMPLER VARIABILITY .....                                             | 4         |
| 2.2.1 Data processing.....                                                                    | 4         |
| 2.3 CALCULATING COMBINED STANDARD UNCERTAINTY .....                                           | 5         |
| 2.4 REMARKS .....                                                                             | 5         |
| <b>3.0 RESULTS.....</b>                                                                       | <b>6</b>  |
| 3.1 DATA SET 1 – BETWEEN LABORATORY AND INTERNAL LABORATORY<br>VARIABILITY .....              | 6         |
| 3.2 DATA SET 2 – BETWEEN SAMPLER VARIABILITY .....                                            | 8         |
| 3.3 DATA SET 3A – SITE SPECIFIC RANKING.....                                                  | 9         |
| 3.4 DATA SET 3B - UNCERTAINTY OVER THE MEASURED CONCENTRATION RANGE                           | 14        |
| 3.4.1 Internal repeatability by individual laboratories .....                                 | 20        |
| 3.5 DETECTION LIMITS .....                                                                    | 21        |
| 3.5.1 Detection limit calculated from sampled filters with low concentrations .....           | 22        |
| 3.6 FIELD BLANKS .....                                                                        | 23        |
| <b>4. CONCLUSIONS.....</b>                                                                    | <b>27</b> |
| <b>ANNEX 1 EQUIVALENCE CALCULATIONS FOR MULTI ANGLE ABSORPTION<br/>PHOTOMETER (MAPP).....</b> | <b>28</b> |
| <b>ANNEX 2 EQUIVALENCE CALCULATIONS FOR SEVEN WAVELENGTH<br/>AETHALOMETER.....</b>            | <b>31</b> |
| <b>ANNEX 3 FILTER CONCENTRATION DATA AND RESULTS OF CALCULATIONS.....</b>                     | <b>35</b> |

## 1. INTRODUCTION

Daily PM<sub>2.5</sub> samples were collected from 6 monitoring stations over 2013 & 2014 and analysed by 4 different laboratories. Sub-samples of each daily filter were analysed twice using 6 thermal-optical protocols (EUSAAR2, NIOSH & IMPROVE) with optical corrections for OCEC split point performed by both reflectance and transmittance methods. This resulted in 12 results per laboratory, per sampled filter. NPL performed a statistical analysis to determine the uncertainty of a single measured value based on between-laboratory variability, internal-laboratory variability and between sampler variability. NPL expanded this analysis to rank the different protocol / optical correction methodologies on a site specific and complete dataset basis.

Laboratory blanks analysed to determine detection limits for OC, EC and TC. Sampled filters with low EC content were also analysed to determine EC detection limits.

OC, EC and TC levels were also determined for field blank samples.

At all the field validation sites two different types of automatic optical black carbon analysers (Multi Angle Absorption Photometer & Aethalometer) were co-located with the filter samplers. Results from these automatic analysers were compared against the filter measurements using the EU Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods<sup>1</sup> and the EU Excel spreadsheet<sup>2</sup>. The results of these equivalence calculations can be seen in Annexes 1 & 2.

## 2. ANALYSIS METHODOLOGY

For the purpose of the data analysis described below, the data are divided into three data sets as follows:

1. Data set 1 relates to those sites for which a single sampler is used by all the laboratories. The sampler used at a particular site may be the same as that used at a different site or it may be different from the samplers used at all other sites. The data set is used to investigate between- and within-laboratory effects and to quantify those effects.
2. Data set 2 relates to those sites for which at least two samplers are used by all the laboratories. As for data set 1, a sampler used at a particular site may be the same as that used at a different site, or it may be different from the samplers used at all other sites. The data set is used to investigate a between-sampler effect and to quantify that effect.
3. Data sets 1 & 2 were combined and split into:
  - a. Sites to determine specific site classification (rural background, urban background & roadside) uncertainties.
  - b. Four concentration ranges to determine uncertainty over the range of concentrations.

The objective of the data analysis is to evaluate the standard uncertainty to be associated with an individual measured value of concentration, which combines the within-laboratory, between-laboratory and between-sampler effects. For data sets 3a and 3b only the within-laboratory and between-laboratory effects were calculated.

### 2.1 CALCULATING BETWEEN- AND WITHIN-LABORATORY VARIABILITY

#### Notation

Let there be  $S$  sites identified by the index  $i, i \in \{1, \dots, S\}$ , and  $L$  laboratories identified by the index:  $j, j \in \{1, \dots, L\}$ . Suppose measurements are made by all  $L$  laboratories at site  $i$  on each of the days identified by the index  $k, k \in K_i$ . The measured values of concentration are then denoted by:

---

1 Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods, Report by an EC Working Group on Guidance for the Demonstration of Equivalence, January 2010.

2 Ruben Beijl, Orthogonal regression and equivalence test utility, V2.9, RIVM, January 2010.

$$x_{ijk}, \quad k \in K_i, \quad j \in \{1, \dots, L\}, \quad i \in \{1, \dots, S\}.$$

The data set is balanced in the sense that all laboratories make measurements at all sites. However, the days  $K_i$  on which measurements are made can be different from one site to another.

### 2.1.1 Data processing

The data processing comprises two stages. In a first stage, the  $L$  measured values corresponding to each site and each day are processed to remove outlying values, and normalised to remove effects associated with the factors of site, sampler and time. In a second stage, an ANalysis Of VAriance (ANOVA) is applied to the resulting data set corresponding to all sites and all days to decide whether a between-laboratory effect exists. If the effect exists, a calculation of the between- and within-laboratory standard deviations is undertaken. If the effect does not exist, a calculation of the within-laboratory standard deviation only is made. In either case, the within-laboratory standard deviation describes the repeatability standard deviation for the laboratories, which is assumed to be the same for all laboratories.

Each part of the data processing is described below.

### 2.1.2 Outlier rejection

Outlier rejection is applied to the set of measured values  $x_{ijk}, j \in \{1, \dots, L\}$ , provided by the  $L$  laboratories on day  $k$  at site  $i$ . Outlier rejection is applied separately for the sets of measured values provided on different days and at different sites.

Outlier rejection involves the following steps:

1. Remove any measured value that is zero, with zero interpreted as denoting a missing value;
2. For the non-zero measured values, evaluate the modified Z-score defined by

$$z_{ijk} = \frac{x_{ijk} - \text{med}(\{x_{ijk}\})}{1.483 \times \text{mad}(\{x_{ijk}\})},$$

where:

$\text{med}(\{x_{ijk}\}) = \hat{x}_{ik}$  is the median of the non-zero values  $\{x_{ijk} : x_{ijk} \neq 0, j = 1, \dots, L\}$ , a robust measure of location, and  $\text{mad}(\{x_{ijk}\}) = \text{med}(\{|x_{ijk} - \hat{x}_{ik}|\})$  is the median-absolute deviation of those values, a robust measure of dispersion;

3. Remove any measured value for which  $|z_{ijk}| > 3.5$ .

The procedure for outlier rejection described in steps 2 and 3 follows the recommendation made in section 1.3.5.17 of the NIST Engineering Statistics Handbook<sup>3</sup>. Outlier rejection can be problematic for small sample sizes, in this case when the number  $L$  of laboratories is small. The critical value (of, here, 3.5) for the absolute value of a modified Z-score, which is used to decide whether a measured value is to be considered as an outlier or not, can be adjusted to make outlier rejection easier (using a smaller value of the critical value) or harder (using a larger value). However, such an adjustment leads to a degree of “subjectivity” into the step of outlier rejection, and ought to be justified.

For the majority of the data analysis performed for this standard outlier rejection was not used in order to calculate the uncertainty of a single measured value.

---

<sup>3</sup> [www.itl.nist.gov/div898/handbook/](http://www.itl.nist.gov/div898/handbook/)

### 2.1.3 Data normalisation

Following outlier rejection, the data for site  $i$  and day  $k$  is denoted by:

$$x_{ijk}, \quad k \in K_i, \quad j \in J_{ik}, \quad i \in \{1, \dots, S\},$$

where  $J_{ik}$  contains the indices of the laboratories for which measured values are retained.

Data normalisation involves the following steps:

1. Evaluate the average

$$\bar{x}_{ik} = \frac{1}{L_{ik}} \sum_{j \in J_{ik}} x_{ijk},$$

where  $L_{ik}$  is the number of indices in  $J_{ik}$ ;

2. Shift the measured values by the average  $\bar{x}_{ik}$ , and scale the shifted values by  $\bar{x}_{ik}$  to obtain normalised values

$$v_{ijk} = \frac{x_{ijk} - \bar{x}_{ik}}{\bar{x}_{ik}}, \quad j \in J_{ik}.$$

The average value calculated in step 1 is taken as a reference or consensus value for the concentration of OC, EC or TC at the specific site  $i$  on the specific day  $k$ . The aim of shifting the measured values is to remove, at least approximately, any dependence of measured concentration on the time of measurement, the site at which the measurement is made, and the sampler that is used to make the measurement. The aim of scaling the shifted values is to remove, at least approximately, any dependence of the variability of measured concentration on the value of concentration. The application of scaling is based on the assumption that the repeatability standard deviation is proportional to the measured concentration, i.e., the relative repeatability standard deviation is approximately constant. The normalised data is reported as a fraction (or percentage) of the reference value.

### 2.1.4 Analysis of variance

Following outlier rejection and data normalisation, the data is denoted by:

$$v_{ijk}, \quad k \in K_i, \quad j \in J_{ik}, \quad i \in \{1, \dots, S\}.$$

Since there is no interest in the factors of site and time, an equivalent representation of the data is as

$$y_{lr}, \quad r \in \{1, \dots, R_l\}, \quad l \in \{1, \dots, L\},$$

which groups the measured values by laboratory. Here,  $R_l$  is the number of retained measured values for laboratory  $l$  over all sites and days, and each  $y_{lr}$  equates to one of the normalised values  $v_{ijk}$ .

An ANOVA is used to test the null hypothesis that the averages for the laboratories are equal, i.e., there is no laboratory effect. The ANOVA calculation must account for the fact that the data set may be unbalanced, because the numbers of measured values can be different from one laboratory to another. The function “anovan”, which is provided in Matlab’s Statistics Toolbox<sup>4</sup>, can be used to perform a (multiway) ANOVA and allows for an unbalanced data set. The function returns a  $p$  value that is compared with a critical value  $p_c$ , usually 0.01 or 0.05, chosen before the analysis. If  $p \geq p_c$ , the null hypothesis of no laboratory effect is

---

<sup>4</sup> Matlab®, The MathWorks Inc.

accepted at a  $100p_c\%$  level of confidence. Otherwise, the null hypothesis is rejected at that level of confidence.

### 2.1.5 Calculation of standard deviations

The within-laboratory variance (squared standard deviation) is calculated as the pooled variance  $s^2$  of the laboratory values, and is given by

$$s^2 = \frac{1}{(\sum_{l=1}^L R_l) - L} \sum_{l=1}^L (R_l - 1) s_l^2,$$

where

$$\bar{y}_l = \frac{1}{R_l} \sum_{r=1}^{R_l} y_{lr}, \quad s_l^2 = \frac{1}{R_l - 1} \sum_{r=1}^{R_l} (y_{lr} - \bar{y}_l)^2.$$

If the ANOVA indicates the existence of a between-laboratory effect, then that effect is quantified by the standard deviation  $s_{\text{lab}}$  of the averages calculated for the laboratories, i.e.,

$$s_{\text{lab}}^2 = \frac{1}{L - 1} \sum_{l=1}^L (\bar{y}_l - \bar{\bar{y}})^2,$$

where

$$\bar{\bar{y}} = \frac{1}{L} \sum_{l=1}^L \bar{y}_l.$$

In the case that the null hypothesis of no laboratory effect is accepted, the between-laboratory standard deviation is taken to be zero.

## 2.2 CALCULATING BETWEEN-SAMPLER VARIABILITY

### Notation

Let there be  $S$  sites identified by the index  $i, i \in \{1, \dots, S\}$ , and  $L$  laboratories identified by the index  $j, j \in \{1, \dots, L\}$ . Suppose measurements are made by all  $L$  laboratories at site  $i$  on each of the days identified by the index  $k, k \in K_i$ , using samplers identified by index  $f, f \in F_i$ . The measured values of concentration are then denoted by:

$$x_{ijkf}, \quad f \in F_i, \quad k \in K_i, \quad j \in \{1, \dots, L\}, \quad i \in \{1, \dots, S\}.$$

The data set is balanced in the sense that all laboratories make measurements at all sites. However, the days  $K_i$  on which measurements are made, as well as the samplers  $F_i$  used, can be different from one site to another.

### 2.2.1 Data processing

In a similar way to that for data set 1, the data processing comprises two stages. In a first stage, considering the data for each sampler separately, the measured values corresponding to each site and each day are processed to remove outlying values. Then, considering the data for the samplers together, the values corresponding to each site and each day are normalised to remove effects associated with the factors of site and time, but preserving any sampler effect. In a second stage, an ANOVA is applied to the resulting data set corresponding to all sites, all days and all laboratories to decide whether a between-sampler effect exists. If the effect exists, a calculation of the between-sampler standard deviations is undertaken.

Following outlier rejection and data normalisation, the data is denoted by:

$$v_{ijkf}, \quad f \in F_i, \quad k \in K_i, \quad j \in J_{ikf}, \quad i \in \{1, \dots, S\},$$

where  $J_{ikf}$  contains the indices of the laboratories for which measured values are retained for site  $i$ , day  $k$  and sampler  $f$ . Since there is no interest in the factors of site, time and laboratory, an equivalent representation of the data is as:

$$y_{fr}, \quad r \in \{1, \dots, R_f\}, \quad f \in F_1 \cup \dots \cup F_S,$$

which groups the measured values by sampler. Here,  $R_f$  is the number of retained measured values for sampler  $f$  over all sites, days and laboratories, and each  $y_{fr}$  equates to one of the normalised values  $v_{ijkf}$ . As for data set 1, the data set is generally unbalanced, because the numbers of measured values can be different from one sampler to another.

If an ANOVA indicates the existence of a between-sampler effect, then that effect is quantified by the standard deviation  $s_{\text{sam}}$  of the averages calculated for the samplers, i.e.,

$$s_{\text{sam}}^2 = \frac{1}{F-1} \sum_{f=1}^F (\bar{y}_f - \bar{y})^2,$$

where

$$\bar{y}_f = \frac{1}{R_f} \sum_{r=1}^{R_f} y_{fr}, \quad \bar{y} = \frac{1}{F} \sum_{f=1}^F \bar{y}_f,$$

and  $F$  is the total number of samplers. In the case that the null hypothesis of no sampler effect is accepted, the between-sampler standard deviation is taken to be zero.

## 2.3 CALCULATING COMBINED STANDARD UNCERTAINTY

The relative standard uncertainty  $u_{\text{rel}}$  associated with an individual measured value of concentration is given:

$$u_{\text{rel}}^2 = s^2 + s_{\text{lab}}^2 + s_{\text{sam}}^2,$$

which combines the standard deviations related to within-laboratory, between-laboratory and between-sampler effects.

The between sampler variability results showed that  $s_{\text{sam}}^2$  was insignificant and was not included in the uncertainty calculation.

## 2.4 REMARKS

The data analysis described, including the use of ANOVA and the calculations of the standard deviations to quantify the various effects, depends on assumptions about the homogeneity of these effects for different concentrations, sites, laboratories, samplers and time. Graphical displays of the data, in which the data is plotted against these factors, can be useful to identify obvious departures from these assumptions. In cases that the assumptions do not hold, the results of the data analysis may not be reliable expressions of the various effects considered.


### 3.0 RESULTS

All results were calculated with the outlier rejection disabled.

Spreadsheets containing all the input data and the results of all calculations can be viewed in Annex 3.

#### 3.1 DATA SET 1 – BETWEEN LABORATORY AND INTERNAL LABORATORY VARIABILITY

The following 2 figures show the graphical output from the analysis for the determination of between laboratory variability. The first figure shows a clear between laboratory variance and the second figure shows minimal between laboratory variability.



Each colour represents a different laboratory's result. There are 8 data sets as repeat analyses were considered to be separate data sets to allow the calculation of the uncertainty from a single measured value.

The following table gives the  $p$  values from the analysis of variance for between laboratory effects. If  $p \geq p_c$ , the null hypothesis of no laboratory effect is accepted at a  $100p_c$  % level of confidence. Otherwise, the null hypothesis is rejected at that level of confidence.

| Protocol      | EC       | OC       | TC       |
|---------------|----------|----------|----------|
| EUSAAR2 Trans | 3.00E-04 | 0.0006   | 0.0003   |
| IMPROVE Trans | 1.61E-05 | 1.07E-09 | 8.80E-17 |
| NIOSH Trans   | 8.92E-29 | 0.2399   | 1.20E-18 |
| EUSAAR2 Refl  | 0.0003   | 0.0012   | 0.0003   |
| IMPROVE Refl  | 0.0519   | 3.03E-09 | 8.65E-17 |
| NIOSH Refl    | 0.3544   | 2.35E-38 | 1.19E-18 |

#### $p$ Value Results from Between Laboratory Analysis of Variance

It can be seen that only IMPROVE reflectance (EC), NIOSH reflectance (EC) and NIOSH transmittance (OC) show no significant between laboratory effect at the 5% level of confidence. In all cases the between laboratory variability was calculated and included in the overall uncertainty calculations.

Between laboratory variability and internal laboratory variability were calculated for the 6 thermal / optical protocols and the results shown below. Protocols were ranked: 0.75, 2, 3, 4, 5 & 6 with a ranking of 0.75 denoting the lowest result, rising to a ranking of 6 for the highest result. Rankings for between laboratory variability and internal laboratory variability were combined to provide an overall performance rank for the protocols.

| Score | Analysis Protocol | Component | Between laboratory variability, % |
|-------|-------------------|-----------|-----------------------------------|
| 0.75  | EUSAAR2 Trans     | EC        | 2.2                               |
| 2     | IMPROVE Trans     | EC        | 2.9                               |
| 3     | NIOSH Trans       | EC        | 9.0                               |
| 4     | EUSAAR2 Refl      | EC        | 12.9                              |
| 5     | IMPROVE Refl      | EC        | 67.2                              |
| 6     | NIOSH Refl        | EC        | 488.7                             |
|       |                   |           |                                   |
| 0.75  | NIOSH Trans       | OC        | 0.6                               |
| 2     | NIOSH Refl        | OC        | 3.3                               |
| 3     | IMPROVE Trans     | OC        | 2.1                               |
| 4     | EUSAAR2 Trans     | OC        | 1.5                               |
| 5     | EUSAAR2 Refl      | OC        | 1.5                               |
| 6     | IMPROVE Refl      | OC        | 2.3                               |
|       |                   |           |                                   |
| 0.75  | NIOSH Refl        | TC        | 1.6                               |
| 2     | NIOSH Trans       | TC        | 1.6                               |
| 3     | EUSAAR2 Trans     | TC        | 1.3                               |
| 4     | EUSAAR2 Refl      | TC        | 1.3                               |
| 5     | IMPROVE Trans     | TC        | 2.4                               |
| 6     | IMPROVE Refl      | TC        | 2.4                               |

#### Between Laboratory Variability

| Score | Analysis Protocol | Component | Internal laboratory variability, % |
|-------|-------------------|-----------|------------------------------------|
| 0.75  | EUSAAR2 Trans     | EC        | 9.4                                |
| 2     | IMPROVE Trans     | EC        | 11.5                               |
| 3     | NIOSH Trans       | EC        | 16.1                               |
| 4     | EUSAAR2 Refl      | EC        | 55.6                               |
| 5     | IMPROVE Refl      | EC        | 399.2                              |
| 6     | NIOSH Refl        | EC        | 3922.5                             |
|       |                   |           |                                    |
| 0.75  | NIOSH Trans       | OC        | 4.3                                |
| 2     | EUSAAR2 Trans     | OC        | 6.7                                |
| 3     | EUSAAR2 Refl      | OC        | 7.1                                |
| 4     | IMPROVE Trans     | OC        | 6.4                                |
| 5     | IMPROVE Refl      | OC        | 7.0                                |
| 6     | NIOSH Refl        | OC        | 5.1                                |
|       |                   |           |                                    |
| 0.75  | EUSAAR2 Trans     | TC        | 5.9                                |
| 2     | EUSAAR2 Refl      | TC        | 5.9                                |
| 3     | NIOSH Refl        | TC        | 3.5                                |
| 4     | NIOSH Trans       | TC        | 3.5                                |
| 5     | IMPROVE Trans     | TC        | 5.6                                |
| 6     | IMPROVE Refl      | TC        | 5.6                                |

#### Internal Laboratory Variability

| EC            | Between laboratory | Internal laboratory | Combined Rank |
|---------------|--------------------|---------------------|---------------|
| EUSAAR2 Trans | 0.75               | 0.75                | 1.5           |
| IMPROVE Trans | 2                  | 2                   | 4             |
| NIOSH Trans   | 3                  | 3                   | 6             |
| EUSAAR2 Refl  | 4                  | 4                   | 8             |
| IMPROVE Refl  | 5                  | 5                   | 10            |
| NIOSH Refl    | 6                  | 6                   | 12            |
|               |                    |                     |               |
| OC            | Between laboratory | Internal laboratory | Combined Rank |
| NIOSH Trans   | 0.75               | 0.75                | 1.5           |
| EUSAAR2 Trans | 2                  | 4                   | 6             |
| IMPROVE Trans | 4                  | 3                   | 7             |
| NIOSH Refl    | 6                  | 2                   | 8             |
| EUSAAR2 Refl  | 3                  | 6                   | 9             |
| IMPROVE Refl  | 5                  | 5                   | 10            |
|               |                    |                     |               |
| TC            | Between laboratory | Internal laboratory | Combined Rank |
| NIOSH Refl    | 3                  | 0.75                | 3.75          |
| NIOSH Trans   | 4                  | 2                   | 6             |
| EUSAAR2 Trans | 0.75               | 6                   | 6.75          |
| EUSAAR2 Refl  | 2                  | 5                   | 7             |
| IMPROVE Trans | 5                  | 4                   | 9             |
| IMPROVE Refl  | 6                  | 3                   | 9             |

### Combined between laboratory and internal laboratory variability protocol ranking

## 3.2 DATA SET 2 – BETWEEN SAMPLER VARIABILITY

The following summarises the results of the between sampler variability:

- No significant between sampler variability was found for TC.
- One protocol gave a possibly significant between sampler variability for OC. IMPROVE Trans, 5.5% level, resulting in an inter-sampler variability of 1.94% ( $2\sigma$ , 95%).
- Three protocols showed a significant to very significant between sampler variability for EC. EUSAAR2 Trans, 3.2% level, IMPROVE Trans, 0.3% level & NIOSH Trans 0.5% level, resulting in a between sampler variability of 5.8, 6.0 and 8.8% ( $2\sigma$ , 95%) respectively. This significant between sampler variability may have only been detected due to the lower repeatability of these 3 protocols, 18%, 22% & 32% ( $2\sigma$ , 95%).
- The other 3 protocols, EUSAAR2 Refl, IMPROVE Refl & NIOSH Refl had internal repeatability's of 112%, 800% and 7800% ( $2\sigma$ , 95%) respectively making the determination of between sampler repeatability impossible.
- There was no significant between sampler variability detected for most components and most protocols. Where it was detected it was insignificant compared to the internal repeatability.

Between sampler variability was not included in the overall uncertainty calculations.

### 3.3 DATA SET 3A – SITE SPECIFIC RANKING

Combined variability was calculated from the between laboratory variability and internal laboratory variability and the results are shown below for EC, OC and TC:

| Score | Analysis Protocol | Component | Site | Combined variability, % |
|-------|-------------------|-----------|------|-------------------------|
| 0.75  | IMPROVE Trans     | EC        | CAB  | 13.6                    |
| 2     | EUSAAR2 Trans     | EC        | CAB  | 19.0                    |
| 3     | NIOSH Trans       | EC        | CAB  | 23.5                    |
| 4     | IMPROVE Refl      | EC        | CAB  | 27.3                    |
| 5     | EUSAAR2 Refl      | EC        | CAB  | 28.7                    |
| 6     | NIOSH Refl        | EC        | CAB  | 34.2                    |
|       |                   |           |      |                         |
| 0.75  | IMPROVE Refl      | EC        | IPR  | 10.8                    |
| 2     | IMPROVE Trans     | EC        | IPR  | 11.0                    |
| 3     | NIOSH Refl        | EC        | IPR  | 11.4                    |
| 4     | EUSAAR2 Refl      | EC        | IPR  | 15.8                    |
| 5     | EUSAAR2 Trans     | EC        | IPR  | 15.9                    |
| 6     | NIOSH Trans       | EC        | IPR  | 16.3                    |
|       |                   |           |      |                         |
| 0.75  | EUSAAR2 Trans     | EC        | WAL  | 15.3                    |
| 2     | IMPROVE Trans     | EC        | WAL  | 19.7                    |
| 3     | NIOSH Trans       | EC        | WAL  | 30.1                    |
| 4     | EUSAAR2 Refl      | EC        | WAL  | 117.7                   |
| 5     | IMPROVE Refl      | EC        | WAL  | 880.0                   |
| 6     | NIOSH Refl        | EC        | WAL  | 8573.5                  |
|       |                   |           |      |                         |
| 0.75  | IMPROVE Trans     | EC        | BCN  | 4.8                     |
| 2     | EUSAAR2 Trans     | EC        | BCN  | 6.6                     |
| 3     | NIOSH Trans       | EC        | BCN  | 12.3                    |
| 4     | EUSAAR2 Refl      | EC        | BCN  | 16.3                    |
| 5     | IMPROVE Refl      | EC        | BCN  | 19.1                    |
| 6     | NIOSH Refl        | EC        | BCN  | 30.1                    |
|       |                   |           |      |                         |
| 0.75  | EUSAAR2 Trans     | EC        | DUI  | 5.2                     |
| 2     | NIOSH Refl        | EC        | DUI  | 6.6                     |
| 3     | IMPROVE Trans     | EC        | DUI  | 8.9                     |
| 4     | EUSAAR2 Refl      | EC        | DUI  | 10.1                    |
| 5     | IMPROVE Refl      | EC        | DUI  | 11.3                    |
| 6     | NIOSH Trans       | EC        | DUI  | 15.0                    |
|       |                   |           |      |                         |
| 0.75  | EUSAAR2 Trans     | EC        | AMS  | 9.3                     |
| 2     | IMPROVE Trans     | EC        | AMS  | 10.1                    |
| 3     | EUSAAR2 Refl      | EC        | AMS  | 11.9                    |
| 4     | NIOSH Trans       | EC        | AMS  | 13.0                    |
| 5     | NIOSH Refl        | EC        | AMS  | 17.1                    |
| 6     | IMPROVE Refl      | EC        | AMS  | 17.3                    |

| Score | Analysis Protocol | Component | Site | Combined variability, % |
|-------|-------------------|-----------|------|-------------------------|
| 0.75  | EUSAAR2 Trans     | OC        | CAB  | 7.1                     |
| 2     | IMPROVE Trans     | OC        | CAB  | 7.4                     |
| 3     | NIOSH Trans       | OC        | CAB  | 9.2                     |
| 4     | NIOSH Refl        | OC        | CAB  | 11.9                    |
| 5     | EUSAAR2 Refl      | OC        | CAB  | 12.5                    |
| 6     | IMPROVE Refl      | OC        | CAB  | 14.3                    |
|       |                   |           |      |                         |
| 0.75  | NIOSH Trans       | OC        | IPR  | 5.1                     |
| 2     | IMPROVE Trans     | OC        | IPR  | 5.4                     |
| 3     | NIOSH Refl        | OC        | IPR  | 5.7                     |
| 4     | IMPROVE Refl      | OC        | IPR  | 7.2                     |
| 5     | EUSAAR2 Trans     | OC        | IPR  | 11.3                    |
| 6     | EUSAAR2 Refl      | OC        | IPR  | 11.5                    |
|       |                   |           |      |                         |
| 0.75  | NIOSH Trans       | OC        | WAL  | 5.0                     |
| 2     | NIOSH Refl        | OC        | WAL  | 5.5                     |
| 3     | EUSAAR2 Refl      | OC        | WAL  | 9.1                     |
| 4     | EUSAAR2 Trans     | OC        | WAL  | 9.6                     |
| 5     | IMPROVE Refl      | OC        | WAL  | 11.3                    |
| 6     | IMPROVE Trans     | OC        | WAL  | 11.6                    |
|       |                   |           |      |                         |
| 0.75  | NIOSH Trans       | OC        | BCN  | 3.3                     |
| 2     | IMPROVE Trans     | OC        | BCN  | 4.0                     |
| 3     | EUSAAR2 Trans     | OC        | BCN  | 5.5                     |
| 4     | EUSAAR2 Refl      | OC        | BCN  | 6.9                     |
| 5     | IMPROVE Refl      | OC        | BCN  | 7.1                     |
| 6     | NIOSH Refl        | OC        | BCN  | 9.4                     |
|       |                   |           |      |                         |
| 0.75  | EUSAAR2 Trans     | OC        | DUI  | 2.9                     |
| 2     | NIOSH Trans       | OC        | DUI  | 3.5                     |
| 3     | NIOSH Refl        | OC        | DUI  | 3.6                     |
| 4     | IMPROVE Trans     | OC        | DUI  | 4.4                     |
| 5     | EUSAAR2 Refl      | OC        | DUI  | 4.7                     |
| 6     | IMPROVE Refl      | OC        | DUI  | 5.3                     |
|       |                   |           |      |                         |
| 0.75  | IMPROVE Refl      | OC        | AMS  | 5.0                     |
| 2     | IMPROVE Trans     | OC        | AMS  | 5.0                     |
| 3     | NIOSH Trans       | OC        | AMS  | 5.5                     |
| 4     | NIOSH Refl        | OC        | AMS  | 5.8                     |
| 5     | EUSAAR2 Trans     | OC        | AMS  | 8.2                     |
| 6     | EUSAAR2 Refl      | OC        | AMS  | 8.3                     |

| Score | Analysis Protocol | Component | Site | Combined variability, % |
|-------|-------------------|-----------|------|-------------------------|
| 0.75  | EUSAAR2 Refl      | TC        | CAB  | 6.0                     |
| 2     | EUSAAR2 Trans     | TC        | CAB  | 6.0                     |
| 3     | IMPROVE Refl      | TC        | CAB  | 6.5                     |
| 4     | IMPROVE Trans     | TC        | CAB  | 6.5                     |
| 5     | NIOSH Refl        | TC        | CAB  | 8.4                     |
| 6     | NIOSH Trans       | TC        | CAB  | 8.4                     |
|       |                   |           |      |                         |
| 0.75  | IMPROVE Trans     | TC        | IPR  | 4.3                     |
| 2     | IMPROVE Refl      | TC        | IPR  | 4.3                     |
| 3     | NIOSH Refl        | TC        | IPR  | 4.6                     |
| 4     | NIOSH Trans       | TC        | IPR  | 4.6                     |
| 5     | EUSAAR2 Refl      | TC        | IPR  | 11.3                    |
| 6     | EUSAAR2 Trans     | TC        | IPR  | 11.3                    |
|       |                   |           |      |                         |
| 0.75  | NIOSH Refl        | TC        | WAL  | 5.2                     |
| 2     | NIOSH Trans       | TC        | WAL  | 5.2                     |
| 3     | EUSAAR2 Refl      | TC        | WAL  | 8.9                     |
| 4     | EUSAAR2 Trans     | TC        | WAL  | 8.9                     |
| 5     | IMPROVE Refl      | TC        | WAL  | 11.0                    |
| 6     | IMPROVE Trans     | TC        | WAL  | 11.0                    |
|       |                   |           |      |                         |
| 0.75  | NIOSH Refl        | TC        | BCN  | 3.0                     |
| 2     | NIOSH Trans       | TC        | BCN  | 3.0                     |
| 3     | IMPROVE Refl      | TC        | BCN  | 3.3                     |
| 4     | IMPROVE Trans     | TC        | BCN  | 3.3                     |
| 5     | EUSAAR2 Trans     | TC        | BCN  | 4.2                     |
| 6     | EUSAAR2 Refl      | TC        | BCN  | 4.2                     |
|       |                   |           |      |                         |
| 0.75  | EUSAAR2 Trans     | TC        | DUI  | 2.7                     |
| 2     | EUSAAR2 Refl      | TC        | DUI  | 2.7                     |
| 3     | NIOSH Refl        | TC        | DUI  | 3.2                     |
| 4     | NIOSH Trans       | TC        | DUI  | 3.2                     |
| 5     | IMPROVE Trans     | TC        | DUI  | 4.1                     |
| 6     | IMPROVE Refl      | TC        | DUI  | 4.1                     |
|       |                   |           |      |                         |
| 0.75  | IMPROVE Refl      | TC        | AMS  | 3.4                     |
| 2     | IMPROVE Trans     | TC        | AMS  | 3.4                     |
| 3     | NIOSH Refl        | TC        | AMS  | 4.2                     |
| 4     | NIOSH Trans       | TC        | AMS  | 4.2                     |
| 5     | EUSAAR2 Refl      | TC        | AMS  | 6.8                     |
| 6     | EUSAAR2 Trans     | TC        | AMS  | 6.8                     |

The results for EC, OC and TC were combined by site classification to give an overall ranking per site classification as shown below:

### Rural Background

| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 4.75        | IMPROVE Trans | 4.5         | NIOSH Trans   | 8.75        | EUSAAR2 Refl  |
| 7.75        | EUSAAR2 Trans | 9           | NIOSH Refl    | 8.75        | NIOSH Refl    |
| 9.75        | IMPROVE Refl  | 9.75        | EUSAAR2 Trans | 10          | IMPROVE Refl  |
| 12          | NIOSH Trans   | 10          | IMPROVE Trans | 10.75       | IMPROVE Trans |
| 13          | EUSAAR2 Refl  | 14          | EUSAAR2 Refl  | 12          | EUSAAR2 Trans |
| 15          | NIOSH Refl    | 15          | IMPROVE Refl  | 12          | NIOSH Trans   |

### Urban Background

| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 2.75        | EUSAAR2 Trans | 2.75        | NIOSH Trans   | 3.75        | NIOSH Refl    |
| 3.75        | IMPROVE Trans | 3.75        | EUSAAR2 Trans | 5.75        | EUSAAR2 Trans |
| 8           | EUSAAR2 Refl  | 6           | IMPROVE Trans | 6           | NIOSH Trans   |
| 8           | NIOSH Refl    | 9           | EUSAAR2 Refl  | 8           | EUSAAR2 Refl  |
| 9           | NIOSH Trans   | 9           | NIOSH Refl    | 9           | IMPROVE Refl  |
| 10          | IMPROVE Refl  | 11          | IMPROVE Refl  | 9           | IMPROVE Trans |

### Roadside

| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 0.75        | EUSAAR2 Trans | 0.75        | IMPROVE Refl  | 0.75        | IMPROVE Refl  |
| 2           | IMPROVE Trans | 2           | IMPROVE Trans | 2           | IMPROVE Trans |
| 3           | EUSAAR2 Refl  | 3           | NIOSH Trans   | 3           | NIOSH Refl    |
| 4           | NIOSH Trans   | 4           | NIOSH Refl    | 4           | NIOSH Trans   |
| 5           | NIOSH Refl    | 5           | EUSAAR2 Trans | 5           | EUSAAR2 Refl  |
| 6           | IMPROVE Refl  | 6           | EUSAAR2 Refl  | 6           | EUSAAR2 Trans |

This ranking was repeated just taking internal laboratory variability into account to try and remove any between laboratory dependence. The results are shown below:

**Rural Background**

| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 4.75        | IMPROVE Trans | 7.5         | NIOSH Trans   | 6.5         | IMPROVE Refl  |
| 7.75        | EUSAAR2 Trans | 8.75        | IMPROVE Trans | 9.75        | NIOSH Refl    |
| 8.75        | IMPROVE Refl  | 9           | NIOSH Refl    | 10          | IMPROVE Trans |
| 11          | NIOSH Trans   | 11          | EUSAAR2 Refl  | 11          | EUSAAR2 Refl  |
| 15          | EUSAAR2 Refl  | 12          | EUSAAR2 Trans | 11          | NIOSH Trans   |
| 15          | NIOSH Refl    | 14          | IMPROVE Refl  | 14          | EUSAAR2 Trans |

**Urban Background**


| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 3.75        | IMPROVE Trans | 3.75        | NIOSH Trans   | 3.75        | IMPROVE Refl  |
| 3.75        | EUSAAR2 Trans | 4           | IMPROVE Trans | 5.75        | NIOSH Refl    |
| 4           | NIOSH Refl    | 4.75        | EUSAAR2 Trans | 6           | IMPROVE Trans |
| 9           | EUSAAR2 Refl  | 7           | NIOSH Refl    | 8           | NIOSH Trans   |
| 10          | IMPROVE Refl  | 11          | IMPROVE Refl  | 9           | EUSAAR2 Refl  |
| 11          | NIOSH Trans   | 11          | EUSAAR2 Refl  | 9           | EUSAAR2 Trans |


**Roadside**

| EC          |               | OC          |               | TC          |               |
|-------------|---------------|-------------|---------------|-------------|---------------|
| Total score | Protocol      | Total score | Protocol      | Total score | Protocol      |
| 0.75        | IMPROVE Trans | 0.75        | IMPROVE Trans | 0.75        | IMPROVE Refl  |
| 2           | EUSAAR2 Trans | 2           | IMPROVE Refl  | 2           | IMPROVE Trans |
| 3           | NIOSH Trans   | 3           | NIOSH Trans   | 3           | NIOSH Refl    |
| 4           | EUSAAR2 Refl  | 4           | NIOSH Refl    | 4           | NIOSH Trans   |
| 5           | NIOSH Refl    | 5           | EUSAAR2 Trans | 5           | EUSAAR2 Refl  |
| 6           | IMPROVE Refl  | 6           | EUSAAR2 Refl  | 6           | EUSAAR2 Trans |

### 3.4 DATA SET 3B - UNCERTAINTY OVER THE MEASURED CONCENTRATION RANGE

Results for each thermal / optical protocol from the complete data set were ordered by concentration and then split into 4 concentration bins: 0% - 25%, 25% - 50%, 50% - 75% & 75% - 100% with equal number of measurements in each concentration bin. The combined expanded uncertainty ( $2\sigma$ , 95%) due to between laboratory variability and internal laboratory variability were calculated for each concentration bin, the results of which are shown below:





The following tables give the full results:


| Analysis Protocol | Component | Concentration quarter | Between laboratory repeatability, % | Internal laboratory repeatability, % | Expanded combined Uncertainty, % | Minimum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Maximum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Mid-range, $\mu\text{g}\cdot\text{cm}^{-2}$ |
|-------------------|-----------|-----------------------|-------------------------------------|--------------------------------------|----------------------------------|-------------------------------------------|-------------------------------------------|---------------------------------------------|
| EUSAAR2 Refl      | EC        | Q1                    | 16.2                                | 86.5                                 | 176.0                            | 0.00                                      | 2.05                                      | 1.03                                        |
| EUSAAR2 Refl      | EC        | Q2                    | 2.1                                 | 8.3                                  | 17.1                             | 2.08                                      | 3.97                                      | 3.03                                        |
| EUSAAR2 Refl      | EC        | Q3                    | 4.3                                 | 9.6                                  | 21.1                             | 3.97                                      | 6.86                                      | 5.42                                        |
| EUSAAR2 Refl      | EC        | Q4                    | 6.6                                 | 8.7                                  | 21.8                             | 6.87                                      | 22.57                                     | 14.72                                       |
| EUSAAR2 Trans     | EC        | Q1                    | 2.6                                 | 13.7                                 | 27.8                             | 0.00                                      | 1.54                                      | 0.77                                        |
| EUSAAR2 Trans     | EC        | Q2                    | 4.5                                 | 6.8                                  | 16.4                             | 1.55                                      | 2.52                                      | 2.04                                        |
| EUSAAR2 Trans     | EC        | Q3                    | 4.0                                 | 7.6                                  | 17.2                             | 2.53                                      | 4.38                                      | 3.46                                        |
| EUSAAR2 Trans     | EC        | Q4                    | 4.2                                 | 7.1                                  | 16.6                             | 4.39                                      | 15.49                                     | 9.94                                        |
| IMPROVE Refl      | EC        | Q1                    | 98.0                                | 619.0                                | 1253.0                           | 0.00                                      | 1.62                                      | 0.81                                        |
| IMPROVE Refl      | EC        | Q2                    | 4.5                                 | 7.8                                  | 18.1                             | 1.63                                      | 3.37                                      | 2.50                                        |
| IMPROVE Refl      | EC        | Q3                    | 4.3                                 | 7.7                                  | 17.6                             | 3.38                                      | 5.98                                      | 4.68                                        |
| IMPROVE Refl      | EC        | Q4                    | 9.2                                 | 9.1                                  | 25.8                             | 5.98                                      | 18.43                                     | 12.21                                       |
| IMPROVE Trans     | EC        | Q1                    | 1.9                                 | 16.0                                 | 32.3                             | 0.00                                      | 1.82                                      | 0.91                                        |
| IMPROVE Trans     | EC        | Q2                    | 3.2                                 | 6.5                                  | 14.5                             | 1.82                                      | 2.95                                      | 2.39                                        |
| IMPROVE Trans     | EC        | Q3                    | 4.0                                 | 6.4                                  | 15.1                             | 2.95                                      | 4.76                                      | 3.86                                        |
| IMPROVE Trans     | EC        | Q4                    | 3.9                                 | 7.3                                  | 16.5                             | 4.78                                      | 18.96                                     | 11.87                                       |
| NIOSH Refl        | EC        | Q1                    | 741.0                               | 6064.0                               | 12218.0                          | 0.00                                      | 1.62                                      | 0.81                                        |
| NIOSH Refl        | EC        | Q2                    | 2.7                                 | 9.3                                  | 19.4                             | 1.63                                      | 3.37                                      | 2.50                                        |
| NIOSH Refl        | EC        | Q3                    | 4.2                                 | 9.0                                  | 19.9                             | 3.38                                      | 5.98                                      | 4.68                                        |
| NIOSH Refl        | EC        | Q4                    | 5.7                                 | 9.5                                  | 22.2                             | 5.98                                      | 18.43                                     | 12.21                                       |
| NIOSH Trans       | EC        | Q1                    | 7.2                                 | 23.1                                 | 48.4                             | 0.00                                      | 1.11                                      | 0.56                                        |
| NIOSH Trans       | EC        | Q2                    | 5.0                                 | 8.9                                  | 20.4                             | 1.11                                      | 1.95                                      | 1.53                                        |
| NIOSH Trans       | EC        | Q3                    | 5.7                                 | 9.8                                  | 22.7                             | 1.96                                      | 3.38                                      | 2.67                                        |
| NIOSH Trans       | EC        | Q4                    | 6.6                                 | 9.4                                  | 22.9                             | 3.38                                      | 9.28                                      | 6.33                                        |


| Analysis Protocol | Component | Concentration quarter | Between laboratory repeatability, % | Internal laboratory repeatability, % | Expanded combined Uncertainty, % | Minimum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Maximum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Mid-range, $\mu\text{g}\cdot\text{cm}^{-2}$ |
|-------------------|-----------|-----------------------|-------------------------------------|--------------------------------------|----------------------------------|-------------------------------------------|-------------------------------------------|---------------------------------------------|
| EUSAAR2 Refl      | OC        | Q1                    | 2.4                                 | 7.9                                  | 16.6                             | 2.46                                      | 6.62                                      | 4.54                                        |
| EUSAAR2 Refl      | OC        | Q2                    | 0.7                                 | 4.0                                  | 8.1                              | 6.62                                      | 8.57                                      | 7.60                                        |
| EUSAAR2 Refl      | OC        | Q3                    | 1.3                                 | 5.3                                  | 10.8                             | 8.59                                      | 12.79                                     | 10.69                                       |
| EUSAAR2 Refl      | OC        | Q4                    | 2.4                                 | 5.3                                  | 11.6                             | 12.82                                     | 40.24                                     | 26.53                                       |
| EUSAAR2 Trans     | OC        | Q1                    | 1.8                                 | 6.8                                  | 14.0                             | 2.62                                      | 7.03                                      | 4.83                                        |
| EUSAAR2 Trans     | OC        | Q2                    | 0.9                                 | 3.6                                  | 7.5                              | 7.04                                      | 9.81                                      | 8.43                                        |
| EUSAAR2 Trans     | OC        | Q3                    | 1.3                                 | 4.6                                  | 9.5                              | 9.81                                      | 15.05                                     | 12.43                                       |
| EUSAAR2 Trans     | OC        | Q4                    | 0.9                                 | 4.0                                  | 8.2                              | 15.09                                     | 45.33                                     | 30.21                                       |
| IMPROVE Refl      | OC        | Q1                    | 4.4                                 | 8.8                                  | 19.7                             | 2.91                                      | 7.13                                      | 5.02                                        |
| IMPROVE Refl      | OC        | Q2                    | 0.9                                 | 4.6                                  | 9.4                              | 7.47                                      | 9.23                                      | 8.35                                        |
| IMPROVE Refl      | OC        | Q3                    | 1.0                                 | 4.7                                  | 9.7                              | 9.24                                      | 13.63                                     | 11.44                                       |
| IMPROVE Refl      | OC        | Q4                    | 4.0                                 | 5.9                                  | 14.2                             | 13.63                                     | 44.10                                     | 28.87                                       |
| IMPROVE Trans     | OC        | Q1                    | 2.6                                 | 5.7                                  | 12.5                             | 2.46                                      | 6.83                                      | 4.65                                        |
| IMPROVE Trans     | OC        | Q2                    | 1.7                                 | 3.8                                  | 8.4                              | 6.83                                      | 9.51                                      | 8.17                                        |
| IMPROVE Trans     | OC        | Q3                    | 2.0                                 | 4.0                                  | 8.9                              | 9.51                                      | 14.56                                     | 12.04                                       |
| IMPROVE Trans     | OC        | Q4                    | 1.1                                 | 4.2                                  | 8.7                              | 14.57                                     | 43.88                                     | 29.23                                       |
| NIOSH Refl        | OC        | Q1                    | 2.6                                 | 6.9                                  | 14.7                             | 2.91                                      | 6.54                                      | 4.73                                        |
| NIOSH Refl        | OC        | Q2                    | 0.5                                 | 3.4                                  | 6.9                              | 7.47                                      | 9.23                                      | 8.35                                        |
| NIOSH Refl        | OC        | Q3                    | 0.7                                 | 5.2                                  | 10.4                             | 9.24                                      | 13.63                                     | 11.44                                       |
| NIOSH Refl        | OC        | Q4                    | 2.2                                 | 5.6                                  | 12.0                             | 13.64                                     | 44.10                                     | 28.87                                       |
| NIOSH Trans       | OC        | Q1                    | 1.5                                 | 5.3                                  | 11.1                             | 2.75                                      | 7.74                                      | 5.25                                        |
| NIOSH Trans       | OC        | Q2                    | 0.7                                 | 3.5                                  | 7.1                              | 7.75                                      | 10.69                                     | 9.22                                        |
| NIOSH Trans       | OC        | Q3                    | 1.3                                 | 3.9                                  | 8.3                              | 10.69                                     | 15.97                                     | 13.33                                       |
| NIOSH Trans       | OC        | Q4                    | 1.0                                 | 3.6                                  | 7.4                              | 15.99                                     | 52.09                                     | 34.04                                       |

| Analysis Protocol | Component | Concentration quarter | Between laboratory repeatability, % | Internal laboratory repeatability, % | Expanded combined Uncertainty, % | Minimum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Maximum, $\mu\text{g}\cdot\text{cm}^{-2}$ | Mid-range, $\mu\text{g}\cdot\text{cm}^{-2}$ |
|-------------------|-----------|-----------------------|-------------------------------------|--------------------------------------|----------------------------------|-------------------------------------------|-------------------------------------------|---------------------------------------------|
| EUSAAR2 Refl      | TC        | Q1                    | 2.2                                 | 5.7                                  | 12.2                             | 3.01                                      | 8.74                                      | 5.88                                        |
| EUSAAR2 Refl      | TC        | Q2                    | 1.2                                 | 3.6                                  | 7.7                              | 8.76                                      | 12.67                                     | 10.72                                       |
| EUSAAR2 Refl      | TC        | Q3                    | 1.2                                 | 3.5                                  | 7.4                              | 12.68                                     | 18.44                                     | 15.56                                       |
| EUSAAR2 Refl      | TC        | Q4                    | 1.5                                 | 3.3                                  | 7.2                              | 18.45                                     | 57.90                                     | 38.18                                       |
| EUSAAR2 Trans     | TC        | Q1                    | 2.2                                 | 5.7                                  | 12.2                             | 3.01                                      | 8.74                                      | 5.88                                        |
| EUSAAR2 Trans     | TC        | Q2                    | 1.2                                 | 3.6                                  | 7.7                              | 8.76                                      | 12.67                                     | 10.72                                       |
| EUSAAR2 Trans     | TC        | Q3                    | 1.2                                 | 3.5                                  | 7.4                              | 12.38                                     | 18.44                                     | 15.41                                       |
| EUSAAR2 Trans     | TC        | Q4                    | 1.5                                 | 3.3                                  | 7.2                              | 18.45                                     | 57.90                                     | 38.18                                       |
| IMPROVE Refl      | TC        | Q1                    | 2.3                                 | 4.7                                  | 10.6                             | 3.14                                      | 9.09                                      | 6.12                                        |
| IMPROVE Refl      | TC        | Q2                    | 1.9                                 | 3.4                                  | 7.7                              | 9.10                                      | 12.16                                     | 10.63                                       |
| IMPROVE Refl      | TC        | Q3                    | 1.9                                 | 3.0                                  | 7.1                              | 12.92                                     | 18.80                                     | 15.86                                       |
| IMPROVE Refl      | TC        | Q4                    | 1.6                                 | 3.2                                  | 7.1                              | 18.82                                     | 59.68                                     | 39.25                                       |
| IMPROVE Trans     | TC        | Q1                    | 2.4                                 | 4.7                                  | 10.6                             | 3.10                                      | 8.92                                      | 6.01                                        |
| IMPROVE Trans     | TC        | Q2                    | 1.8                                 | 3.4                                  | 7.7                              | 8.92                                      | 12.81                                     | 10.86                                       |
| IMPROVE Trans     | TC        | Q3                    | 1.9                                 | 3.0                                  | 7.1                              | 12.82                                     | 18.48                                     | 15.65                                       |
| IMPROVE Trans     | TC        | Q4                    | 1.6                                 | 3.2                                  | 7.1                              | 18.51                                     | 59.13                                     | 38.82                                       |
| NIOSH Refl        | TC        | Q1                    | 1.4                                 | 4.8                                  | 10.0                             | 3.14                                      | 9.09                                      | 6.12                                        |
| NIOSH Refl        | TC        | Q2                    | 1.5                                 | 3.0                                  | 6.7                              | 9.10                                      | 12.91                                     | 11.01                                       |
| NIOSH Refl        | TC        | Q3                    | 1.4                                 | 3.2                                  | 6.9                              | 12.92                                     | 18.80                                     | 15.86                                       |
| NIOSH Refl        | TC        | Q4                    | 1.5                                 | 3.1                                  | 6.8                              | 18.82                                     | 59.68                                     | 39.25                                       |
| NIOSH Trans       | TC        | Q1                    | 1.4                                 | 4.8                                  | 10.0                             | 3.14                                      | 9.09                                      | 6.12                                        |
| NIOSH Trans       | TC        | Q2                    | 1.5                                 | 3.0                                  | 6.7                              | 9.10                                      | 12.91                                     | 11.01                                       |
| NIOSH Trans       | TC        | Q3                    | 1.4                                 | 3.2                                  | 6.9                              | 11.04                                     | 18.80                                     | 14.92                                       |
| NIOSH Trans       | TC        | Q4                    | 1.5                                 | 3.1                                  | 6.8                              | 18.82                                     | 59.68                                     | 39.25                                       |

As the above uncertainties are calculated on a single measurement result, it shows that repeat analysis is not required to deliver an overall uncertainty of less than 25% for concentrations above  $2 \mu\text{g}\cdot\text{cm}^{-2}$ . For concentrations below this the relative uncertainty is greater than 25%, but probably limited by the detection limit.

The above results show that the combined uncertainty for the reflectance optical correction increases dramatically at low concentrations due to very large internal repeatabilities. This is due to the reflectance method being unable to detect small EC concentrations. This is highlighted by the charts below:





### 3.4.1 Internal repeatability by individual laboratories

In addition to analysing sub samples each analysis laboratory performed repeated measurements from a single filter to get a separate measure of internal repeatability. The results are presented in the following table:

| Protocol          | Site | OC Repeatability,<br>% | EC Repeatability,<br>% | TC Repeatability,<br>% |
|-------------------|------|------------------------|------------------------|------------------------|
| <b>ERL</b>        |      |                        |                        |                        |
| No protocol Trans | IPR  | 7.3                    | 17.6                   | 5.7                    |
| No protocol Refl  | IPR  | 8.1                    | 11.3                   | 5.7                    |
| <b>GGD</b>        |      |                        |                        |                        |
| No protocol Trans | IPR  | 4.1                    | 12.9                   | 5.7                    |
| <b>IUTA</b>       |      |                        |                        |                        |
| EUSAAR2 Trans     | AMS  | 6.9                    | 10.9                   | 6.4                    |
| IMPROVE Trans     | AMS  | 6.9                    | 12.9                   | 5.4                    |
| NIOSH Trans       | AMS  | 8.2                    | 25.7                   | 4.9                    |
| EUSAAR2 Trans     | CAB  | 12.3                   | 28.9                   | 3.2                    |
| IMPROVE Trans     | CAB  | 8.1                    | 18.7                   | 8.4                    |
| NIOSH Trans       | CAB  | 7.7                    | 12.5                   | 7.8                    |
| EUSAAR2 Trans     | IPR  | 7.3                    | 10.8                   | 6.9                    |

The above results are similar to the overall uncertainties calculated in Section 3.4.

### 3.5 DETECTION LIMITS

Laboratory blank filters were repeatedly analysed by each laboratory and the results pooled to determine the method detection limit in  $\mu\text{g.cm}^{-2}$ , for each protocol according to the following equation:

$$\text{Detection Limit} = SD * t_{99,n-1}$$

Where:

- $SD$  = the standard deviation of the repeated measurements,  $\mu\text{g.cm}^{-2}$
- $T_{99, n-1}$  = the student t factor with a level of confidence of 99% for  $n-1$  degrees of freedom
- $n$  = the number of repeat measurements

The following tables give the calculated method detection limits for OC, EC and TC for each protocol:

|         |       | n  | SD, $\mu\text{g.cm}^{-2}$ | DL, $\mu\text{g.cm}^{-2}$ | Mean, $\mu\text{g.cm}^{-2}$ |
|---------|-------|----|---------------------------|---------------------------|-----------------------------|
| EUSAAR2 | Refl  | 14 | 0.731                     | 1.852                     | 1.364                       |
|         | Trans | 14 | 0.691                     | 1.750                     | 1.326                       |
| IMPROVE | Refl  | 14 | 0.878                     | 2.224                     | 1.455                       |
|         | Trans | 14 | 0.843                     | 2.136                     | 1.388                       |
| NIOSH   | Refl  | 12 | 1.087                     | 2.818                     | 2.162                       |
|         | Trans | 12 | 0.902                     | 2.338                     | 2.040                       |

#### OC Detection Limit

The proposed standard specifies that the maximum OC content of a laboratory blank filter shall be  $\leq 2.0 \mu\text{g.cm}^{-2}$ . It appears that the NIOSH protocols are over reporting OC as the same filters were analysed by the other 2 protocols. This OC level is generally higher than that seen in other measurement studies.

|         |       | <b>n</b> | <b>SD, <math>\mu\text{g.cm}^{-2}</math></b> | <b>DL, <math>\mu\text{g.cm}^{-2}</math></b> | <b>Mean, <math>\mu\text{g.cm}^{-2}</math></b> |
|---------|-------|----------|---------------------------------------------|---------------------------------------------|-----------------------------------------------|
| EUSAAR2 | Refl  | 14       | 0.007                                       | 0.017                                       | -0.002                                        |
|         | Trans | 14       | 0.071                                       | 0.179                                       | 0.036                                         |
| IMPROVE | Refl  | 14       | 0.000                                       | 0.000                                       | 0.000                                         |
|         | Trans | 14       | 0.093                                       | 0.237                                       | 0.066                                         |
| NIOSH   | Refl  | 12       | 0.001                                       | 0.002                                       | 0.000                                         |
|         | Trans | 12       | 0.256                                       | 0.663                                       | 0.122                                         |

#### EC Detection Limit

The mean detection limit for reflectance protocols is  $<0.1 \mu\text{g.cm}^{-2}$ , while the mean detection limit for transmittance protocols was  $\sim 0.4 \mu\text{g.cm}^{-2}$ . This was above the proposals for the draft standard and is higher than that seen in other measurement studies.

|         |       | <b>n</b> | <b>SD, <math>\mu\text{g.cm}^{-2}</math></b> | <b>DL, <math>\mu\text{g.cm}^{-2}</math></b> | <b>Mean, <math>\mu\text{g.cm}^{-2}</math></b> |
|---------|-------|----------|---------------------------------------------|---------------------------------------------|-----------------------------------------------|
| EUSAAR2 | Refl  | 14       | 0.734                                       | 1.858                                       | 1.362                                         |
|         | Trans | 14       | 0.734                                       | 1.858                                       | 1.362                                         |
| IMPROVE | Refl  | 14       | 0.878                                       | 2.224                                       | 1.455                                         |
|         | Trans | 14       | 0.878                                       | 2.224                                       | 1.455                                         |
| NIOSH   | Refl  | 12       | 1.087                                       | 2.818                                       | 2.162                                         |
|         | Trans | 12       | 1.087                                       | 2.818                                       | 2.162                                         |

#### TC Detection Limit

### 3.5.1 Detection limit calculated from sampled filters with low concentrations

The detection limit was also calculated using the same equation as in Section 3.5 from sampled filters with low concentrations. Filters with a negative or zero content were removed from the data set before calculation. There were no sampled filters with an OC or TC content  $<2.0 \mu\text{g.cm}^{-2}$ , therefore detection limits for OC and TC have not been calculated.

For filters with low EC content a cut-off threshold of  $0.4 \mu\text{g.cm}^{-2}$  was chosen as this represents twice the acceptable laboratory filter blank limit. The resulting detection limits for each protocol are shown in the following table:

|         |       | <b>n</b> | <b>SD, <math>\mu\text{g.cm}^{-2}</math></b> | <b>DL, <math>\mu\text{g.cm}^{-2}</math></b> | <b>Mean, <math>\mu\text{g.cm}^{-2}</math></b> | <b>No. of filters excluded</b> |
|---------|-------|----------|---------------------------------------------|---------------------------------------------|-----------------------------------------------|--------------------------------|
| EUSAAR2 | Refl  | 21       | 0.101                                       | 0.245                                       | 0.053                                         | 110                            |
|         | Trans | 59       | 0.044                                       | 0.102                                       | 0.331                                         | 2                              |
| IMPROVE | Refl  | 39       | 0.045                                       | 0.105                                       | 0.012                                         | 95                             |
|         | Trans | 29       | 0.040                                       | 0.096                                       | 0.355                                         | 2                              |
| NIOSH   | Refl  | 42       | 0.097                                       | 0.225                                       | 0.044                                         | 99                             |
|         | Trans | 82       | 0.074                                       | 0.168                                       | 0.269                                         | 5                              |

The difference between the reflectance and transmittance protocols in measuring low levels of EC is best demonstrated by ranked frequency charts as shown in Section 3.4. The validity of the detection

limit results for the transmittance protocols is limited due to the low number of samples below the cut-off threshold.

### **3.6 FIELD BLANKS**

Field blanks were analysed by each laboratory using the 6 protocols. Field blank OC content was similar to the laboratory blank content for all sites for all protocols. Field blank EC content was at or below the detection limit results calculated from the laboratory blanks.

The following tables provide the results in detail, blank levels above  $2 \mu\text{g.cm}^{-2}$  have been highlighted in red:

| Site     | Count | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
|          |       | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 31    | 0.942                        | 2.224                        | 2.160                          | 0.000                        | 0.000                        | 0.000                          | 0.942                        | 2.224                        | 2.160                          |
| BCN      | 8     | 0.204                        | 0.579                        | 1.591                          | 0.056                        | 0.159                        | 0.024                          | 0.224                        | 0.637                        | 1.615                          |
| CAB      | 45    | 0.604                        | 1.402                        | 1.932                          | 0.023                        | 0.053                        | 0.002                          | 0.607                        | 1.409                        | 1.934                          |
| DUI      | 16    | 0.695                        | 1.730                        | 2.766                          | 0.010                        | 0.024                        | 0.004                          | 0.694                        | 1.729                        | 2.770                          |
| IPR      | 10    | 0.514                        | 1.380                        | 1.827                          | 0.006                        | 0.017                        | 0.002                          | 0.511                        | 1.372                        | 1.829                          |
| WAL      | 16    | 0.533                        | 1.327                        | 1.583                          | 0.000                        | 0.001                        | 0.000                          | 0.533                        | 1.328                        | 1.583                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 126   | 0.757                        | 1.717                        | 2.020                          | 0.020                        | 0.046                        | 0.003                          | 0.757                        | 1.719                        | 2.023                          |

**Eusaar2 Reflectance**

| Site     | Count | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
|          |       | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 32    | 0.923                        | 2.175                        | 2.060                          | 0.120                        | 0.282                        | 0.083                          | 0.932                        | 2.195                        | 2.144                          |
| BCN      | 16    | 0.212                        | 0.529                        | 1.626                          | 0.014                        | 0.036                        | 0.005                          | 0.215                        | 0.536                        | 1.631                          |
| CAB      | 45    | 0.613                        | 1.423                        | 1.926                          | 0.051                        | 0.117                        | 0.012                          | 0.608                        | 1.411                        | 1.938                          |
| DUI      | 16    | 0.694                        | 1.729                        | 2.770                          | 0.000                        | 0.000                        | 0.000                          | 0.694                        | 1.729                        | 2.770                          |
| IPR      | 14    | 0.473                        | 1.198                        | 1.830                          | 0.038                        | 0.097                        | 0.025                          | 0.473                        | 1.198                        | 1.855                          |
| WAL      | 16    | 0.488                        | 1.216                        | 1.558                          | 0.065                        | 0.162                        | 0.025                          | 0.533                        | 1.328                        | 1.583                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 139   | 0.726                        | 1.644                        | 1.969                          | 0.075                        | 0.170                        | 0.029                          | 0.730                        | 1.655                        | 1.998                          |

**Eusaar2 Transmittance**

| Site     | Count | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
|          |       | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 8     | 0.449                        | 1.276                        | 1.973                          | 0.000                        | 0.000                        | 0.000                          | 0.449                        | 1.276                        | 1.973                          |
| BCN      | 8     | 0.365                        | 1.038                        | 1.402                          | 0.429                        | 1.220                        | 0.253                          | 0.203                        | 0.577                        | 1.655                          |
| CAB      | 43    | 0.459                        | 1.068                        | 1.890                          | 0.140                        | 0.325                        | 0.042                          | 0.466                        | 1.083                        | 1.932                          |
| DUI      | 16    | 0.774                        | 1.927                        | 2.748                          | 0.207                        | 0.515                        | 0.128                          | 0.756                        | 1.883                        | 2.876                          |
| IPR      | 10    | 0.542                        | 1.454                        | 1.501                          | 0.020                        | 0.053                        | -0.005                         | 0.545                        | 1.463                        | 1.496                          |
| WAL      | 16    | 0.696                        | 1.733                        | 1.669                          | 0.000                        | 0.000                        | 0.000                          | 0.696                        | 1.733                        | 1.669                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 101   | 0.679                        | 1.544                        | 1.920                          | 0.181                        | 0.411                        | 0.058                          | 0.686                        | 1.561                        | 1.978                          |

**IMPROVE Reflectance**

| Site     | Count | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
|          |       | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 8     | 0.449                        | 1.276                        | 1.973                          | 0.000                        | 0.000                        | 0.000                          | 0.449                        | 1.276                        | 1.973                          |
| BCN      | 16    | 0.179                        | 0.445                        | 1.673                          | 0.000                        | 0.001                        | 0.000                          | 0.179                        | 0.445                        | 1.673                          |
| CAB      | 43    | 0.469                        | 1.090                        | 1.910                          | 0.061                        | 0.141                        | 0.022                          | 0.466                        | 1.083                        | 1.932                          |
| DUI      | 16    | 0.750                        | 1.866                        | 2.856                          | 0.003                        | 0.007                        | 0.101                          | 0.038                        | 0.094                        | 0.344                          |
| IPR      | 14    | 0.650                        | 1.647                        | 1.643                          | 0.134                        | 0.338                        | 0.090                          | 0.692                        | 1.751                        | 1.733                          |
| WAL      | 16    | 0.678                        | 1.688                        | 1.611                          | 0.090                        | 0.224                        | 0.058                          | 0.696                        | 1.733                        | 1.669                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 113   | 0.665                        | 1.511                        | 1.939                          | 0.076                        | 0.173                        | 0.031                          | 0.668                        | 1.517                        | 1.970                          |

**IMPROVE Transmittance**

|          |       | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
| Site     | Count | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 9     | 0.345                        | 0.950                        | 1.679                          | 0.000                        | 0.000                        | 0.000                          | 0.345                        | 0.950                        | 1.679                          |
| BCN      | 8     | 0.812                        | 2.306                        | 2.024                          | 0.031                        | 0.089                        | 0.003                          | 0.820                        | 2.331                        | 2.026                          |
| CAB      | 33    | 0.593                        | 1.394                        | 2.116                          | 0.017                        | 0.040                        | -0.009                         | 0.593                        | 1.395                        | 2.107                          |
| DUI      | 8     | 0.577                        | 1.639                        | 2.810                          | 0.031                        | 0.087                        | 0.009                          | 0.558                        | 1.585                        | 2.820                          |
| IPR      | 8     | 1.050                        | 2.983                        | 2.199                          | 0.000                        | 0.000                        | 0.000                          | 1.050                        | 2.983                        | 2.199                          |
| WAL      | 16    | 0.856                        | 2.133                        | 1.772                          | 0.000                        | 0.000                        | 0.000                          | 0.856                        | 2.133                        | 1.772                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 82    | 0.747                        | 1.707                        | 2.068                          | 0.018                        | 0.041                        | -0.003                         | 0.748                        | 1.708                        | 2.065                          |

**NIOSH Reflectance**

|          |       | OC                           |                              |                                | EC                           |                              |                                | TC                           |                              |                                |
|----------|-------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|------------------------------|------------------------------|--------------------------------|
| Site     | Count | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ | SD,<br>$\mu\text{g.cm}^{-2}$ | DL,<br>$\mu\text{g.cm}^{-2}$ | Mean,<br>$\mu\text{g.cm}^{-2}$ |
| AMS      | 8     | 0.369                        | 1.048                        | 1.684                          | 0.000                        | 0.000                        | 0.000                          | 0.369                        | 1.048                        | 1.684                          |
| BCN      | 16    | 0.592                        | 1.474                        | 1.863                          | 0.000                        | 0.000                        | 0.000                          | 0.592                        | 1.474                        | 1.863                          |
| CAB      | 33    | 0.594                        | 1.396                        | 2.108                          | 0.014                        | 0.034                        | -0.001                         | 0.593                        | 1.395                        | 2.107                          |
| DUI      | 8     | 0.556                        | 1.581                        | 2.823                          | 0.005                        | 0.014                        | -0.003                         | 0.558                        | 1.585                        | 2.820                          |
| IPR      | 12    | 0.876                        | 2.271                        | 2.045                          | 0.028                        | 0.073                        | 0.004                          | 0.875                        | 2.270                        | 2.048                          |
| WAL      | 15    | 0.747                        | 1.875                        | 1.765                          | 0.188                        | 0.472                        | 0.056                          | 0.862                        | 2.164                        | 1.822                          |
|          |       |                              |                              |                                |                              |                              |                                |                              |                              |                                |
| combined | 92    | 0.692                        | 1.577                        | 2.026                          | 0.078                        | 0.177                        | 0.009                          | 0.709                        | 1.615                        | 2.035                          |

**NIOSH Transmittance**

## 4. CONCLUSIONS

The analysis shows that the reflectance protocols have problems detecting very low levels of elemental carbon and are unable to deliver the detection limits required.

The results from the transmittance protocols show no protocol clearly performing better than the others across all of the site types and concentrations encountered during the field validation tests. EUSAAR2 performs best for EC at urban background and roadside locations and second best at rural locations.

There was no significant between sampler variability detected for most components and most protocols. Where it was detected it was insignificant compared to the internal repeatability.

The uncertainty of a single measurement result (95%,  $2\sigma$ ) for the transmittance protocols is less than 25% above concentrations of  $2 \mu\text{g.cm}^{-2}$ . For concentrations below this the relative uncertainty is greater than 25%, but probably limited by the detection limit. When performing repeat measurements of a single sampled filter the repeatability was similar to the calculated uncertainty of a single measurement result.

The OC detection limit measured from laboratory blank filters was close to or exceeded the requirements of the draft standard, with the NIOSH protocols over reporting OC levels compared to the other protocols. The mean EC detection limit for reflectance protocols is  $<0.1 \mu\text{g.cm}^{-2}$ , while the mean detection limit for transmittance protocols was  $\sim 0.4 \mu\text{g.cm}^{-2}$ . This was above the proposals for the draft standard and is higher than that seen in other measurement studies.

The EC detection limit measured from sampled filters with low EC content was similar for both reflectance and transmittance protocols with mean EC detection limits of  $0.192 \mu\text{g.cm}^{-2}$  and  $0.122 \mu\text{g.cm}^{-2}$  for reflectance and transmittance protocols, respectively.

Field blank OC and EC content was at or below the detection limit results calculated from the laboratory blanks.


## ANNEX 1 EQUIVALENCE CALCULATIONS FOR MULTI ANGLE ABSORPTION PHOTOMETER (MAAP)

Two MAAP analysers also took part in the field campaigns and were run alongside the filter samplers.


The procedure outlined in the EU Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods and the EU Excel spreadsheet were used to calculate the equivalence of the MAAP compared to filter samples analysed using the EUSAAR2 Transmittance Thermal-Optical Protocol. Details of the monitoring data are shown below:

| Site      | Analyser   | Start date | Stop date  | Number of data points |
|-----------|------------|------------|------------|-----------------------|
| Ispra     | #36 & #360 | 12/11/2013 | 21/11/2013 | 9                     |
| Barcelona | #36        | 20/02/2014 | 04/04/2014 | 38                    |
| Amsterdam | #36        | 12/06/2014 | 14/07/2014 | 30                    |
| Duisburg  | #360       | 09/02/2014 | 09/04/2014 | 38                    |
| Waldhof   | #360       | 26/05/2014 | 13/07/2014 | 37                    |
| Cabauw    | #36 & #360 | 02/09/2014 | 13/10/2014 | 20                    |

The following chart shows the comparison of the concentrations measured by the 2 instruments at the Ispra and Cabauw sites. The linear regression is the normal least squares fit calculated by Excel and not the orthogonal regression calculated by the EU spreadsheet.



The following 2 charts show the relationship between the 2 instruments and the filter measurements. Where parallel filter measurements were performed these have been averaged to provide 1 result.



The following table gives the results of the orthogonal regression of the combined data of both analysers versus the filter samples, taking  $10 \mu\text{g.m}^{-3}$  as the Limit Value for calculating the percentage uncertainty.

| Data Set  | Un-calibrated Uncertainty,<br>% |
|-----------|---------------------------------|
| Full      | 190                             |
| Ispra     | 108                             |
| Barcelona | 287                             |
| Duisburg  | 232                             |
| Amsterdam | 234                             |
| Waldhof   | 316                             |
| Cabauw    | 232                             |


The equivalence uncertainties above are unacceptably high so calibration of the MAAP data is required. The following table gives the calibration slopes and intercept as calculated by the EU spreadsheet.

| Data Set | Slope | Intercept |
|----------|-------|-----------|
| Full     | 1.882 | N/A       |
| Full     | 1.961 | -0.083    |

Applying these calibrations to the MAAP data provides the following relative uncertainties to the filter method:

| Data Set  | Slope Only,<br>% | Slope + Intercept,<br>% |
|-----------|------------------|-------------------------|
| All       | 1.7              | 9.5                     |
| Ispra     | 47.5             | 53.5                    |
| Barcelona | 49.6             | 39.7                    |
| Duisburg  | 20.1             | 11.4                    |
| Amsterdam | 16.6             | 7.7                     |
| Waldhof   | 69.1             | 58.7                    |
| Cabauw    | 23.6             | 15.0                    |

Although the relative uncertainty of the full data set using both calibration methods is low, the relative uncertainty at most monitoring sites is large. The following charts show the relationship between the calibrated MAAP data and the filter method:




## ANNEX 2 EQUIVALENCE CALCULATIONS FOR SEVEN WAVELENGTH AETHALOMETER

Three 7 wavelength Aethalometer analysers also took part in the field campaigns and were run alongside the filter samplers. Analyser # 142 failed at the start of the Amsterdam field measurements and was replaced by analyser #197. Analyser #143 did not record data correctly during the Cabauw field measurements so these results have not been included.


The procedure outlined in the EU Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods and the EU Excel spreadsheet were used to calculate the equivalence of the Aethalometer compared to filter samples analysed using the EUSAAR2 Transmittance Thermal-Optical Protocol. Details of the monitoring data are shown below:

| Site      | Analyser    | Start date | Stop date  | Number of data points |
|-----------|-------------|------------|------------|-----------------------|
| Ispra     | #142 & #143 | 12/11/2013 | 21/11/2013 | 9                     |
| Barcelona | #142        | 20/02/2014 | 04/04/2014 | 38                    |
| Amsterdam | #197        | 14/06/2014 | 14/07/2014 | 27                    |
| Duisburg  | #143        | 09/02/2014 | 09/04/2014 | 38                    |
| Waldhof   | #143        | 26/05/2014 | 13/07/2014 | 37                    |
| Cabauw    | #197        | 02/09/2014 | 13/10/2014 | 20                    |

The following chart shows the comparison of the concentrations measured by the 2 instruments at the Ispra site. The linear regression is the normal least squares fit calculated by Excel and not the orthogonal regression calculated by the EU spreadsheet.



The following 3 charts show the relationship between the 2 instruments and the filter measurements. Where parallel filter measurements were performed these have been averaged to provide 1 result.



The following table gives the results of the orthogonal regression of the combined data of all analysers versus the filter samples, taking  $10 \mu\text{g.m}^{-3}$  as the Limit Value for calculating the percentage uncertainty.

| Data Set  | Un-calibrated Uncertainty,<br>% |
|-----------|---------------------------------|
| Full      | 177                             |
| Ispra     | 90                              |
| Barcelona | 245                             |
| Duisburg  | 150                             |
| Amsterdam | 305                             |
| Waldhof   | 283                             |
| Cabauw    | 403                             |



The equivalence uncertainties above are unacceptably high so calibration of the Aethalometer data is required. The following table gives the calibration slopes and intercept as calculated by the EU spreadsheet.

| Data Set | Slope | Intercept |
|----------|-------|-----------|
| Full     | 1.846 | N/A       |
| Full     | 1.885 | -0.042    |

Applying these calibrations to the Aethalometer data provides the following relative uncertainties to the filter method:

| <b>Data Set</b> | <b>Slope Only,<br/>%</b> | <b>Slope + Intercept,<br/>%</b> |
|-----------------|--------------------------|---------------------------------|
| All             | 6.2                      | 10.1                            |
| Ispra           | 55.6                     | 58.5                            |
| Barcelona       | 35.6                     | 31.0                            |
| Duisburg        | 20.4                     | 24.1                            |
| Amsterdam       | 65.8                     | 60.4                            |
| Waldhof         | 53.3                     | 48.1                            |
| Cabauw          | 118.0                    | 111.5                           |

Although the relative uncertainty of the full data set using both calibration methods is low, the relative uncertainty at most monitoring sites is large. The following charts show the relationship between the calibrated Aethalometer data and the filter method:



## **ANNEX 3 FILTER CONCENTRATION DATA AND RESULTS OF CALCULATIONS**

Filter concentration data and uncertainty calculation results are supplied as separate Excel spreadsheets.

### Results of filter concentration data used for uncertainty calculations

OCEC Filters.xlsx

### Results of calculations

Matlab Results.xlsx